

	Analyte	Analytes in air - Reporting limit for 1 week exposure period in µg/m ³	Analytes in air - Reporting limit for 24 hrs exposure period in µg/m ³
1	1,1-DCE	2.48	17.79
2	1-DCE	1.55	10.84
3	1,1-DCA	1.86	13.02
4	c-DCE	1.12	7.82
5	CF	1.08	7.53
6	1,2-DCA	0.80	5.59
7	1,1,1-TCA	1.54	10.80
8	Benz	0.85	5.97
9	CT	1.35	9.44
10	TCE	0.66	3.90
11	1,1,2-TCA	0.55	3.83
12	Toluene	0.48	3.35
13	PCE	0.43	3.03
14	ChBenz	0.69	4.81
15	1,1,1,2-TetCA	0.29	2.05
16	EtBenz	0.29	2.04
17	p-m-Xyl	0.30	2.09
18	1,1,2,2-TetCA	0.26	1.84
19	o-Xyl	0.31	2.15
20	1,3-DCB	0.38	2.65
21	1,4-DCB	0.38	2.69
22	1,2-DCB	0.38	2.67
23	Naphthalene	0.039	0.271
24	Acenaphthylene	0.009	0.063
25	Acenaphthene	0.008	0.056
26	Fluorene	0.008	0.056
27	Phenanthrene	0.003	0.021
28	Anthracene	0.002	0.014
29	Fluoranthene	0.002	0.014
30	Pyrene	0.001	0.006

Analyte	Abbreviation
1,1-Dichloroethylene	1,1-DCE
trans-Dichloroethylene	t-DCE
1,1-Dichloroethane	1,1-DCA
cis-Dichloroethylene	c-DCE
Chloroform	CF
1,2-Dichloroethane	1,2-DCA
1,1,1-Trichloroethane	1,1,1-TCA
Benzene	Benz.
Carbon tetrachloride	CT
Trichloroethylene	TCE
1,1,2-Trichloroethane	1,1,2-TCA
Toluene	Toluene
Tetrachloroethylene	PCE
Chlorobenzene	Ch-Benz.
1,1,1,2-Tetrachloroethane	1,1,1,2-TetCA
Ethyl benzene	Et-Benz.
p and m-Xylene	p-m-Xyl
1,1,2,2-Tetrachloroethane	1,1,2,2-Tet CA
o-Xylene	o-Xyl
1,3-Dichlorobenzene	1,3-DCB
1,4-Dichlorobenzene	1,4-DCB
1,2-Dichlorobenzene	1,2-DCB